



CERTIFICATE OF ANALYSIS

Product: Lymphoprep™, 6x500 mL
Catalogue No.: 1114547

Lot No.: 00117
Internal No: 10190135
Expiry date: 2020-04

Quality Control Results

| Test | Requirement | Result |
|--------------------------------|---------------|--------|
| Density, g/cm ³ | 1.077 ± 0.001 | 1.078 |
| Osmolality, mOsm | 290 ± 15 | 281 |
| Endotoxins, EU/mL | < 1.0 | < 0.3 |
| Sterility test Acc. to Ph.Eur. | Sterile | Pass |

Approved by:

Hege Friesgaard (FRIEHE) 2017-08-04 08:29:53 (YYYY-MM-DD hh:mm:ss)

This document is electronically signed and is valid without a handwritten signature.