



## CERTIFICATE OF ANALYSIS

Product: Lymphoprep™ Tube, 18x10 mL  
Catalogue No.: 1019818  
Lot No.: 10189058  
Expiry date: 2019-02

### Quality Control Results

Test	Requirement	Result
Density, g/cm <sup>3</sup>	1.077 ± 0.001	1.077
Osmolality, mOsm	290 ± 15	281
Endotoxins, EU/mL	< 1.0	<0.13
Sterility test Acc. to Ph.Eur.	Sterile	Pass

### Approved by:

Anita Pettersen (AP) 2017-09-13 15:46:13 (YYYY-MM-DD hh:mm:ss)

This document is electronically signed and is valid without a handwritten signature.