

PRODUCT DESCRIPTION

OptiPrep™ is a ready made, sterile and endotoxin tested solution of Iodixanol, 5,5'-[(2-hydroxy-1-3 propanediyl)-bis(acetylamino)] bis [N,N'-bis(2,3dihydroxypropyl)-2,4,6-triiodo-1,3-benzenecarboxamide], designed for the in vitro isolation of biological particles.

Composition

Iodixanol: 60% (w/v) in water

Physical-chemical characteristics:

Density: 1.320 ± 0.001 g/ml (20°C)

Osmolality: 170 ± 15 mOsm

OPTIPREP™ WORKING AND GRADIENT SOLUTIONS FOR CELLS

An illustrative example of preparing working solutions for cells, based on a cell suspension medium (CSM) containing 0.85% (w/v) NaCl, 10 mM Tricine-NaOH, pH 7.4, is given in Table 1. Table 2 describes the density of gradient solutions prepared by dilution of the working solution with CSM. Hepes-buffered media provide solutions of identical densities and refractive indices. Low concentrations of additives (e.g. 1-5 mM MgCl₂ or 1-2 mM EDTA) can be included without significant effect on the density or osmolality of the solutions.

Table 1
OptiPrep™ Working Solutions for mammalian cells

Cell suspension medium (CSM): 0.85% (w/v) NaCl, 10 mM Tricine-NaOH, pH 7.4 (ρ = 1.006 g/ml)
Toprepare a 40% (w/v) iodixanol Working Solution (ρ = 1.215 g/ml; η = 1.3976): Diluent required: 0.85% (w/v) NaCl, 30 mM Tricine-NaOH, pH 7.4. Mix 2 vol of OptiPrep™ with 1 vol of diluent.
Toprepare a 30% (w/v) iodixanol Working Solution (ρ = 1.163 g/ml; η = 1.3820): Diluent required: 0.85% (w/v) NaCl, 20 mM Tricine-NaOH, pH 7.4. Mix 1 vol of OptiPrep™ with 1 vol of diluent.
Dilute Working Solution with CSM to make gradient solutions (Table 2)

Table 2
Properties of iodixanol-NaCl/Tricine solutions¹

% iodixanol (w/v)	η	ρ (g/ml)	40% iodixanol WS (ml)	CSM (ml)
6.00	1.3444	1.037	0.6	3.4
8.00	1.3475	1.048	0.8	3.2
10.00	1.3507	1.058	1.0	3.0
12.00	1.3538	1.069	1.2	2.8
14.00	1.3569	1.079	1.4	2.6
16.00	1.3601	1.090	1.6	2.4
18.00	1.3632	1.100	1.8	2.2
20.00	1.3663	1.111	2.0	2.0
22.00	1.3694	1.121	2.2	1.8
24.00	1.3726	1.132	2.4	1.6
26.00	1.3757	1.142	2.6	1.4
28.00	1.3788	1.153	2.8	1.2
30.00	1.3820	1.163	3.0	1.0
32.00	1.3851	1.174	3.2	0.8
34.00	1.3882	1.184	3.4	0.6
36.00	1.3914	1.195	3.6	0.4
38.00	1.3945	1.205	3.8	0.2
40.00	1.3976	1.215	4.0	0

¹ WS = working solution; CSM = cell suspension medium, η = refractive index; ρ = density
The osmolality of the gradient solutions is in the range 285-300 mOsm.

Alternatively, isoosmotic solutions can be prepared simply by diluting OptiPrep™ with a buffered 0.85 (w/v) NaCl or a routine culture medium such as DMEM or RPMI. All these media (in the absence of added serum) have densities of 1.006-1.007 g/ml (Table 3). A culture medium containing 10% serum has a slightly higher density at approx 1.009 g/ml.

Table 3
Properties of iodixanol-DMEM solutions¹

% iodixanol (w/v)	η	ρ (g/ml)	40% iodixanol WS (ml)	DMEM(ml)
6.00	1.3448	1.038	0.6	3.4
8.00	1.3479	1.049	0.8	3.2
10.00	1.3510	1.059	1.0	3.0
12.00	1.3541	1.070	1.2	2.8
14.00	1.3572	1.080	1.4	2.6
16.00	1.3603	1.090	1.6	2.4
18.00	1.3633	1.101	1.8	2.2
20.00	1.3664	1.111	2.0	2.0
22.00	1.3695	1.122	2.2	1.8
24.00	1.3726	1.132	2.4	1.6
26.00	1.3757	1.143	2.6	1.4
28.00	1.3788	1.153	2.8	1.2
30.00	1.3819	1.163	3.0	1.0
32.00	1.3850	1.174	3.2	0.8
34.00	1.3881	1.184	3.4	0.6
36.00	1.3912	1.195	3.6	0.4
38.00	1.3943	1.205	3.8	0.2
40.00	1.3974	1.216	4.0	0

¹ WS = working solution (4 vol of OptiPrep™ + 2 vol DMEM), η = refractive index; ρ = density
The osmolality of the gradient solutions is in the range 290-305 mOsm.

OPTIPREP™ WORKING AND GRADIENT SOLUTIONS FOR MAMMALIAN ORGANELLES

Table 4 describes the preparation of a general purpose Working Solution for mammalian organelles and also modifications for mitochondria and nuclei. Other minor variations such as omission of the EDTA or inclusion of 0.1% (v/v) ethanol or 1-5 mM DTT, or use of alternative buffers will not affect the density or osmolality of the solutions.

Table 4
OptiPrep™ Working Solutions for mammalian organelles

Homogenization media (HM): General purpose: 0.25 M sucrose, 1 mM EDTA 10 mM Tris-HCl, pH 7.4 (ρ = 1.030 g/ml) Nuclei: 0.25 M sucrose, 25 mM KCl, 5 mM MgCl ₂ , 20 mM Tris-HCl, pH 7.8 (ρ = 1.033 g/ml) Mitochondria: 4.4% (w/v) mannitol, 1 mM EDTA 10 mM Tris-HCl, pH 7.4 (ρ = 1.015 g/ml)
Toprepare a 50% (w/v) iodixanol Working Solution (WS1-3): Mix 5 vol of OptiPrep™ with 1 vol of one of the following diluents: General purpose: 0.25 M sucrose, 6 mM EDTA, 60 mM Tris-HCl, pH 7.4. (ρ = 1.032 g/ml) Nuclei: 150 mM KCl, 30 mM MgCl ₂ , 120 mM Tris-HCl, pH 7.8 (ρ = 1.012 g/ml) Mitochondria: 4.4% (w/v) mannitol, 6 mM EDTA 60 mM Tris-HCl, pH 7.4 (ρ = 1.017 g/ml)
WS1: ρ = 1.272 g/ml; η = 1.4147 WS2: ρ = 1.269 g/ml; η = 1.4148 WS3: ρ = 1.269 g/ml; η = 1.4139
Dilute WS with appropriate HM to make gradient solutions (Tables 5-7)

Table 5
Properties of iodixanol-sucrose solutions¹

% iodixanol (w/v)	η	ρ (g/ml)	50% iodixanol WS1 (ml)	HM1(ml)
6.00	1.3534	1.059	1.2	8.8
8.00	1.3562	1.069	1.6	8.4
10.00	1.3589	1.079	2.0	8.0
12.00	1.3617	1.088	2.4	7.6
14.00	1.3645	1.098	2.8	7.2
16.00	1.3673	1.107	3.2	6.8
18.00	1.3701	1.117	3.6	6.4
20.00	1.3729	1.127	4.0	6.0
22.00	1.3757	1.136	4.4	5.6
24.00	1.3785	1.146	4.8	5.2
26.00	1.3813	1.156	5.2	4.8
28.00	1.3840	1.165	5.6	4.4
30.00	1.3868	1.175	6.0	4.0
32.00	1.3896	1.185	6.4	3.6
34.00	1.3924	1.194	6.8	3.2
36.00	1.3952	1.204	7.2	2.8
38.00	1.3980	1.214	7.6	2.4
40.00	1.4008	1.223	8.0	2.0
42.00	1.4036	1.233	8.4	1.6
44.00	1.4064	1.243	8.8	1.2
46.00	1.4091	1.252	9.2	0.8
48.00	1.4119	1.262	9.6	0.4
50.00	1.4147	1.272		

¹ WS = working solution, HM = homogenization medium; η = refractive index; ρ = density
The osmolality of the gradient solutions is in the range 295-310 mOsm.

Table 6
Properties of iodixanol-sucrose-KCl-MgCl₂ solutions¹

% iodixanol (w/v)	η	ρ (g/ml)	50% iodixanol WS2 (ml)	HM2(ml)
10.0	1.3604	1.080	2.0	8.0
20.0	1.3740	1.127	4.0	6.0
30.0	1.3876	1.175	6.0	4.0
40.0	1.4012	1.222	8.0	2.0
50.0	1.4148	1.269		

¹ WS = working solution, HM = homogenization medium; η = refractive index; ρ = density
The osmolality of the gradient solutions is in the range 320-360 mOsm.

Table 7
Properties of iodixanol-mannitol solutions¹

% iodixanol (w/v)	η	ρ (g/ml)	50% iodixanol WS3 (ml)	HM3(ml)
10.0	1.3548	1.065	2.0	8.0
20.0	1.3796	1.116	4.0	6.0
30.0	1.3844	1.167	6.0	4.0
40.0	1.3991	1.218	8.0	2.0
50.0	1.4139	1.269		

¹ WS = working solution, HM = homogenization medium; η = refractive index; ρ = density
The osmolality of the gradient solutions is in the range 292-310 mOsm.

OPTIPREP™ WORKING AND GRADIENT SOLUTIONS FOR NON-MAMMALIAN ORGANELLES

Media containing 400-600 mM mannitol or sorbitol are common but for yeast organelles. The relationship between osmolality and concentration of sorbitol solutions in water is given in Table 8.

The general strategy for production of gradient solutions of the correct osmolality is described in Tables 9-11.

Table 8
Density (ρ), refractive index (η) and osmolality (σ) of sorbitol solutions in 10 mM Tris-HCl, pH 7.4

% sorbitol (w/v)	η	ρ (g/ml)	σ (mOsm)
4.40	1.3390	1.015	265
8.75	1.3455	1.029	525
10.50	1.3480	1.035	657
12.25	1.3505	1.041	774
17.50	1.3580	1.059	1200

Table 9
OptiPrep™ Working Solutions (sorbitol) for non-mammalian organelles

<p>A 550 mOsm Working Solution Suspension medium (SMA): 8.75% (w/v) sorbitol, 1 mM EDTA, 10 mM Tris-HCl, pH 7.4</p> <p>Toprepare a 40% (w/v) iodixanol Working Solution (WSA) Diluent required: 12.25% sorbitol, 3 mM EDTA, 30 mM Tris-HCl, pH 7.4. Mix 2 vol of OptiPrep™ with 1 vol of diluent</p> <p>Properties of 40% iodixanol WSA ρ = 1.225 g/ml; η = 1.4020; σ = 545 mOsm.</p> <p>Dilute WSA with SMA to make solutions of lower densities (Table 10)</p>
<p>B 750 mOsm Working Solution Suspension medium (SMB): 12.25% (w/v) sorbitol, 1 mM EDTA, 10 mM Tris-HCl, pH 7.4</p> <p>Toprepare a 40% (w/v) iodixanol Working Solution (WSB) Diluent required: 17.5% sorbitol, 3 mM EDTA, 30 mM Tris-HCl, pH 7.4. Mix 2 vol of OptiPrep™ with 1 vol of diluent</p> <p>Properties of 40% iodixanol WSB ρ = 1.231 g/ml; η = 1.4032; σ = 756 mOsm.</p> <p>Dilute WSB with SMB to make solutions of lower densities (Table 11)</p>

Table 10
Properties of iodixanol-sorbitol gradient solutions (approx 550 mOsm)¹

% iodixanol (w/v)	η	ρ (g/ml)	40% iodixanol WSA (ml)	SMA (ml)
6.00	1.3536	1.059	0.6	3.4
8.00	1.3565	1.068	0.8	3.2
10.00	1.3596	1.078	1.0	3.0
12.00	1.3622	1.088	1.2	2.8
14.00	1.3650	1.098	1.4	2.6
16.00	1.3679	1.108	1.6	2.4
18.00	1.3707	1.117	1.8	2.2
20.00	1.3736	1.127	2.0	2.0
22.00	1.3764	1.137	2.2	1.8
24.00	1.3792	1.147	2.4	1.6
26.00	1.3821	1.157	2.6	1.4
28.00	1.3849	1.166	2.8	1.2
30.00	1.3878	1.176	3.0	1.0
32.00	1.3906	1.186	3.2	0.8
34.00	1.3935	1.196	3.4	0.6
36.00	1.3963	1.205	3.6	0.4
38.00	1.3992	1.215	3.8	0.2
40.00	1.4020	1.225	4.0	0

¹ WS = working solution, SM = suspension medium; η = refractive index; ρ = density
The osmolality of the gradient solutions is in the range 545-560 mOsm.

Table 11
Properties of iodixanol-sorbitol gradient solutions (approx 750 mOsm)¹

% iodixanol (w/v)	η	ρ (g/ml)	40% iodixanol WSB (ml)	SMB (ml)
6.00	1.3584	1.070	0.6	3.4
8.00	1.3610	1.079	0.8	3.2
10.00	1.3637	1.089	1.0	3.0
12.00	1.3663	1.098	1.2	2.8
14.00	1.3689	1.108	1.4	2.6
16.00	1.3716	1.117	1.6	2.4
18.00	1.3742	1.127	1.8	2.2
20.00	1.3769	1.136	2.0	2.0
22.00	1.3795	1.146	2.2	1.8
24.00	1.3821	1.155	2.4	1.6
26.00	1.3848	1.165	2.6	1.4
28.00	1.3874	1.174	2.8	1.2
30.00	1.3900	1.184	3.0	1.0
32.00	1.3927	1.193	3.2	0.8
34.00	1.3953	1.203	3.4	0.6
36.00	1.3979	1.212	3.6	0.4
38.00	1.4006	1.222	3.8	0.2
40.00	1.4032	1.231	4.0	0

¹ WS = working solution, SM = suspension medium; η = refractive index; ρ = density
The osmolality of the gradient solutions is in the range 755-850 mOsm.

For more information see "Solutions for the formation of gradients" in the main Axis-Shield Catalogue "Applications and Products"

STABILITY AND STORAGE

OptiPrep™ is stable for 3 years provided the solution is kept sterile and protected from light. Prolonged exposure to direct sunlight leads to release of iodine from the molecule. This effect is negligible when working with these solutions on a day to day basis. OptiPrep™ should be stored between 4°C - 24°C. **Before use invert the bottles several times to mix the contents.**

ORDERING INFORMATION

OptiPrep™ prod. no 1114542 1 x 250 ml

Manufacturer:

Alere Technologies AS
P.O. Box 6863 Rodeløkka
N-0504 Oslo, Norway
Phone: +47 22 04 20 00
Fax: +47 22 04 20 01
www.axis-shield-density-gradient-media.com

ISO13485 certified company